

Application Sheet

HSV IgM

REF 51126

Instrument: Elisys Quattro

REF 16300

Setting:

HSVM_51126_R4.asy

For all essential product information, please refer to the User Manual of the analyzer and the instructions for use for reagents, controls, calibrators and the Certificate of Analysis. Please refer to the instruction for use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyzer.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

Material Required

Material	REF	Comments
HSV IgM	51126	
Reagent Rack 0 Elisys Quattro	16300/38	NC, PC, CON, SUB, STOP
Sample Rack 20 Elisys Quattro	16300/41	sample collection tubes

Additional Notes

The sample incubation time of the validated application is different compared to the protocol described in the package insert.

OD-values of 3.5 or above are outside the photometer's dynamic range. For this reason they are displayed as 4.0.