

# Application Sheet

## HSV IgM

REF 51126

## Instrument: HumaReader Single plus

REF 18000

### Setting:

HSVM\_51126\_R4

For all essential product information, please refer to the User Manual of the analyzer and the instructions for use for reagents, controls, calibrators and the Certificate of Analysis. Please refer to the instruction for use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyzer.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

### Additional Notes

The criteria for Blank, NC, PC and the ratio PC/NC are checked automatically. If the criteria fail because of invalid PC and NC the reading is stopped and no results will be displayed. Failed validity criteria for blank and PC/NC ratio are flagged only, the reading will be finished and the results will be displayed.