

AMH Control HumaCLIA SR

REF 82850

LOT 22001

Version 01 / 03 - 2022

Target Values / Sollwerte / Valores asignados / Valeurs cibles

🕒 2023-03-18

Link for latest version of target value sheet:

<https://www.human.de/clia/target-value-sheets/>



REF	Assay Parameter Ensayo	Valid from Gültig ab Válido a partir de	Units Einheiten Unidades	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo
	Essai	Valable à partir de RGT Lot	Unité		ET	Intervalle		ET	Intervalle
				Level 1			Level 2		
82150	AMH	21001	ng/ml	1.56	0.20	0.96 - 2.16	6.75	0.89	4.08 - 9.42