


AMH Control HumaCLIA SR

REF 82850

LOT 22002

Version 01 / 02 - 2023

Target Values / Sollwerte / Valores asignados / Valeurs cibles

 2024-01-27

Link for latest version of target value sheet:

<https://www.human.de/clia/target-value-sheets/>



REF	Assay Parameter Ensayo	Valid from Gültig ab Válido a partir de	Units Einheiten Unidades	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo
	Essai	Valable à partir de RGT Lot	Unité	ET	ET	Intervalle	ET	ET	Intervalle
				Level 1			Level 2		
82150	AMH	21001	ng/ml	1.34	0.18	0.81 - 1.87	6.54	0.86	3.97 - 9.11

