

Standard RHEUMATOID FACTORS

Version 1/04 - 2022

REF

11361

LOT

22003



2023 - 10 - 31

Calibration Values / Kalibrationswerte / Valores de calibración / Valeurs de calibration

Device Gerät Instrumento Equipement	Reference material/method Referenzmaterial/-methode Material/método de referencia Matériel/méthode de référence	Unit Einheit Unidad Unité	Assigned Value Zugewiesener Wert Valor asignado Valeur assignée
HUMAN RHEUMATOID FACTORS - Immunoturbidimetry Latex enhanced			
Multipurpose reagent on HumaStar 100 & 200	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	118
		kIU/l	118
Multipurpose reagent on Humalyzer 4000	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	118
		kIU/l	118
Multipurpose reagent for other	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	118
		kIU/l	118
HumaStar System reagent on HumaStar 300SR	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	118
		kIU/l	118
HumaStar System reagent on HumaStar 600	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	105
		kIU/l	105

CA_RF_11361

Rev. 001 | valid of 17.03.2021

INF 1136102 - Version 1/04 - 2022

Human

Human Gesellschaft für Biochemica und Diagnostica mbH
Max-Planck-Ring 21 · 65205 Wiesbaden · Germany
Telefon +49 6122-9988-0 · Telefax +49 612-9988-100 · e-Mail human@human.de