

# Standard RHEUMATOID FACTORS

Version 1/02 - 2023

REF

11361

LOT

23002



2024 - 07 - 31

## Calibration Values / Kalibrationswerte / Valores de calibración / Valeurs de calibration

Device Gerät Instrumento Equipement	Reference material/method Referenzmaterial/-methode Material/método de referencia Matériel/méthode de référence	Unit Einheit Unidad Unité	Assigned Value Zugewiesener Wert Valor asignado Valeur assignée
<b>HUMAN RHEUMATOID FACTORS - Immunoturbidimetry Latex enhanced</b>			
Multipurpose reagent on HumaStar 100 & 200	Adjusted to 1 <sup>st</sup> British Standard NIBSC 0064/2	IU/ml	114
		kIU/l	114
Multipurpose reagent on Humalyzer 4000	Adjusted to 1 <sup>st</sup> British Standard NIBSC 0064/2	IU/ml	114
		kIU/l	114
Multipurpose reagent for other	Adjusted to 1 <sup>st</sup> British Standard NIBSC 0064/2	IU/ml	114
		kIU/l	114
HumaStar System reagent on HumaStar 300SR	Adjusted to 1 <sup>st</sup> British Standard NIBSC 0064/2	IU/ml	114
		kIU/l	114
HumaStar System reagent on HumaStar 600	Adjusted to 1 <sup>st</sup> British Standard NIBSC 0064/2	IU/ml	103
		kIU/l	103

CA\_RF\_11361

Rev. 001 | valid of 17.03.2021

INF 1136102 - Version 1/02 - 2023

# Human

Human Gesellschaft für Biochemica und Diagnostica mbH  
Max-Planck-Ring 21 · 65205 Wiesbaden · Germany  
Telefon +49 6122-9988-0 · Telefax +49 612-9988-100 · e-Mail human@human.de