


# AMH Control HumaCLIA SR

REF 82850

LOT 20001

Version 01 / 05 - 2021

Target Values / Sollwerte / Valores asignados / Valeurs cibles

 2021-12-04

Link for latest version of target value sheet:

<https://www.human.de/clia/target-value-sheets/>



REF	Assay Parameter Ensayo	Valid from Gültig ab Válido a partir de	Units Einheiten Unidades	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo
	Essai	Valable à partir de RGT Lot	Unité	ET	Intervalle	ET	Intervalle		
				Level 1			Level 2		
82150	AMH	20001	ng/ml	0,95	0,12	0,58 - 1,32	5,55	0,73	3,37 - 7,73

