

Immunoassay Multi Control HumaCLIA SR

REF 84850

LOT 19001

Version 04 / 07 - 2021

Target Values / Sollwerte / Valores asignados / Valeurs cibles

2022-08-22

Link for latest version of target value sheet:

<https://www.human.de/clia/target-value-sheets/>



REF	Assay Parameter Ensayo Essai	Valid from Gültig ab Válido a partir de Valable à partir de [RGT] Lot	Units Einheiten Unidades Unité	Target	SD	Range	Target	SD	Range
				Sollwert Valor asignado Cible	SD SD ET	Bereich Bereich Intervalo Intervalle	Sollwert Valor asignado Cible	SD SD ET	Bereich Bereich Intervalo Intervalle
				Level 1			Level 2		
84600	TSH	20001	µIU/ml	4,04	0,53	2,45 - 5,63	32,59	3,03	23,50 - 41,68
84615	T3	20001	ng/ml	1,22	0,16	0,74 - 1,70	4,67	0,61	2,83 - 6,51
84620	ft3	20001	pg/ml	3,48	0,46	2,11 - 4,85	10,98	1,02	7,92 - 14,04
84610	ft4	20001	pg/ml	6,86	0,90	4,16 - 9,56	24,38	2,27	17,58 - 31,18
82120	FSH	20001	mIU/ml	3,98	0,52	2,41 - 5,55	44,71	4,16	32,24 - 57,18
82115	LH	20001	mIU/ml	4,70	0,62	2,85 - 6,55	79,29	7,37	57,17 - 101,41
84605	T4	20001	ng/ml	62,60	5,82	45,13 - 80,07	174,05	16,19	125,49 - 222,61
84625	anti-TG	20001	IU/ml	123,05	11,44	88,72 - 157,38	173,47	16,13	125,07 - 221,87
84630	anti-TPO	20001	IU/ml	19,24	1,79	13,87 - 24,61	27,71	2,58	19,98 - 35,44
83100	AFP	20001	ng/ml	7,44	0,98	4,51 - 10,37	117,45	10,92	84,68 - 150,22
82100	HCG*	21001	mIU/ml	2,96	0,39	1,79 - 4,13	120,85	11,24	87,13 - 154,57
85000	25-OH Vitamin D	21001	ng/ml	31,59	2,94	22,78 - 40,40	49,11	4,57	35,41 - 62,81
85055	Vitamin B12*	21001	pg/ml	413,30	38,44	297,99 - 528,61	974,20	90,60	702,40 - 1246,00

* Conformity assessment pending

