

For additional information, please refer to the Operators Manual of the analyzer and check current instructions for use for reagents, controls, calibrators and tables of assigned/analytical values. Typical performance data can be found in the Verification Report of the HumaClot Pro, accessible via

www.human.de/data/gb/vr/15800.pdf
www.human-de.com/data/gb/vr/15800.pdf

If the performance data are not accessible via internet, they can be obtained free of charge from your local distributor.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may affect performance of this and other assays in use on your system and the resulting assay values. It is the responsibility of the user to validate any modifications and their impact on all assay results. The application sheet lists all combinations of controls and calibrators for use with the reagent and instrument system; other combinations are not validated or supported.

Material Required

Material	REF	Size	On-board position
HEMOSTAT Thromboplastin ^{liquid} :			
[RG] Thromboplastin Reagent liquid	31012	6 x 2 ml	R1-R3 with magnetic stirrer and reducer ring
[CAL] HEMOSTAT Calibrator	35500	4 x 1 ml	C1 in sample cup for calibration
[NaCl] 0.9% Sodium Chloride	-	2 ml	C2 in cup, for dilution of calibrator
[CPN] HEMOSTAT Control Plasma Normal	35001	6 x 1 ml	Sample rack position 01-22 or Position C5-C8 (when using QC-program)
[CPA] HEMOSTAT Control Plasma Abnormal	35002	6 x 1 ml	
Cuvette Ring	15800/10	6 x 10 x 32 pcs	Cuvette Ring Rotor
[WASH] HumaClot Pro Wash Solution	15800/20	15 ml	W1
[CLEAN] HumaClot Pro Cleaner	15800/30	15 ml	W2
Sample Cups (2 x 250 pcs) "Human" or Sample Cups (500 pcs) "Hitachi"	15800/25 17470/59	4 ml 2 ml	- -
Reducer Ring (3 pcs)	15800/536		R1 – R3
Magnetic stirrer (10 pcs) (to be cleaned with Wash Solution; REF 15800/20)	15800/50		
Empty vials (50 x 5 ml)	15800/40	-	optional

Additional Notes

The required controls have to be transferred into appropriate sample cups.

On-Board Stability

Material	Name in test protocol	Listed in the test setting as	Time [h]
[RGT] : HEMOSTAT Thromboplastin Reagent liquid	PT Liquid RGT	Start-Reagent	24
[CPN] HEMOSTAT Control Plasma Normal	-	Load as sample or as QC (when using QC-program)	4
[CPA] HEMOSTAT Control Plasma Abnormal	-	Load as sample or as QC (when using QC-program)	4
[CAL] : HEMOSTAT Calibrator	-	Load as calibrator	6

The stated stability data were established under controlled laboratory conditions. The above-mentioned on-board stability values may deviate due to differences in laboratory environmental conditions.

Reagent Settings

Enter the LOT numbers into the reagent settings.

Reagent Setup	
[REF]	31012
Test	HEMOSTAT Thromboplastin ^{liquid} :
Test Setup	Hemostat PT liq
Reagent Name	Hemostat PT RGT
Position in List	11
Abbreviation	PT-Li
LOT	<u>Please insert Lot number</u>
Vial	5ml-HumGL*

*5ml-HumGL (5ml HUMAN Glass Bottle), 5ml-HumPL (5ml HUMAN Plastic Bottle)

Interference Studies

No interference up to ...			
Lipids	mg/dl	250	spiked normal and pathological plasma
Hemoglobin	mg/dl	125	spiked normal and pathological plasma
Bilirubin	mg/dl	12.5	spiked normal and pathological plasma

Performance Characteristics

Measuring range	
Valid Clotting	9-80 sec

Reference Interval
Reference intervals vary from laboratory to laboratory depending on the population served, technique and reagent lot used. Therefore, each laboratory must establish its own reference intervals or verify them whenever one or more of the mentioned variables are changed.
For more information how to establish reference intervals see CLSI document C28-A3.

Calibration Settings

A new standard curve must be established when changing kit LOT, after major maintenance or service, if indicated by quality control results and when required by laboratory control procedures and/or governmental regulations.

Calibration Settings

Test Hemostat PT-Li	
Field Name	Settings
1 st conversion	Interpolation
Unit conversion	s -> %
Mode: in/out	log -> log
Output Format	xxx.xx
2 nd conversion	INR
MNPT	MNPT***-value appears automatically after calibration (s)
ISI	<u>Please insert ISI**</u>
Auto-Calibration	
Diluent	NaCl
Determination	1
Cup	Human/ Hitachi
Calibration Values	
0	90 %
1	50 %
2	25 %
3	12.5 %
4	0 %
Standard	
Concentration	<u>Please insert concentration (%)*</u>
Name	Hemostat CAL
LOT	<u>Please insert LOT Number</u>
Convers. range	5 % – 150 %

* The LOT-specific calibration value can be found on the table of analytical values in the calibration kit.

**The LOT specific ISI value can be found on the table ISI values of the Thromboplastin reagent kit.

Please note:

1st conversion (→ %) and 2rd conversion (→ INR) are optional and can be switched off by entering “none” into the corresponding field.

If only the 1st conversion (→ %) is switched off, it is required to enter the MNPT*** value manually. For further information please refer to the User Manual.

