


SARS-CoV-2 Ag Rapid Test

Fast and reliable – when it matters most

Your advantages at a glance

- > High sensitivity and specificity
- > Use of sputum possible
- > No instrument needed
- > Result within 15 minutes
- > Test option where no laboratory infrastructure is available



Clinical performance¹

Ct value	Number of samples	2019 nCoV RT-PCR results	SARS-CoV-2 antigen test result as compared to RT-PCR
≤ 30	107	pos	103 / 107 = 96.26%
≤ 36	185	pos	161 / 185 = 87.03%
> 40	228	neg	226 / 228 = 99.12%

Days after onset of symptoms	Number of samples	2019 nCoV RT-PCR results	SARS-CoV-2 antigen test result as compared to RT-PCR
≤ 7	103	pos	98 / 103 = 95.15%
≤ 14	133	pos	122 / 133 = 91.73%
> 14	52	pos	39 / 52 = 75.00%

Ct value

The cycle threshold (Ct) is the relevant signal value. It results from the number of replication cycles required to generate a fluorescence signal. Lower Ct values indicate a higher virus concentration. A Ct value of < 40 is clinically reported as PCR positive.²

SARS-CoV-2 Ag Rapid Test REF: LFA0401-25N

- > Colloidal gold immunoassay
- > Sensitivity: 95.15% (≤7 days after onset of symptoms)
Specificity: 99.12%
- > Sample: Nasopharyngeal & oropharyngeal swab specimens
- > Kit size: 25 tests
- > Storage: 2...30°C
- > Shelf life: 18 months
- > CE marked according to Directive 98/79/EC

Kit components

- > 25 test cards with desiccant in a sealed foil pouch
- > Sample extraction solution
- > 25 extraction tubes
- > 25 nasopharyngeal swabs
- > 25 drippers
- > Extraction tube holder
- > Instruction for use

More information about COVID-19 at

www.human.de/covid-19/

¹ SARS-CoV-2 Ag Rapid Test. Package insert; article number LFA0401-25N

² Sethuraman N. et al. (2020) Interpreting Diagnostic Tests for SARS-CoV-2. JAMA. 2020;323(22):2249-2251

Human

Diagnostics Worldwide