

# Application Sheet

**anti-HBc**

**REF 51260**

**Instrument: Elisys Quattro**

**REF 16300**

**Setting:**

**A-HBC\_51260\_R4**

For all essential product information, please refer to the User Manual of the analyzer and the instructions for use for reagents, controls, calibrators and the Certificate of Analysis. Please refer to the instruction for use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyzer.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

## Material Required

Material	REF	Comments
anti-HBc	51260	
Reagent Rack II (2 Tracks) Elisys Quattro	16300/39	NC, PC, CON, SA, SB, STOP
Sample Rack 20 Elisys Quattro	16300/41	Suitable for sample collection tubes

## Additional Notes

Samples must be pipetted within 5 minutes; otherwise a considerable signal drift will affect the results. It is mandatory to use not more than half a plate in one run, since the Quattro cannot calculate results from a second set of controls as specified in the IFU for manual processing.