HSV IgM

[REF] 51126

[REF] **16670**

HSVM 51126 R4

Instrument: HumaReader HS Setting:

For all essential product information, please refer to the User Manual of the analyzer and the instructions for use for reagents, controls, calibrators and the Certificate of Analysis. Please refer to the instruction for use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyzer.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

Additional Notes

The setting requires the definition of at least one well as "CR" (cut-off control), although the assay does not contain a "CR" reagent. Normally the measurement can be started without "CR", but in rare cases you may be asked for it. If this is the case, simply position the marker to an empty well.

The OD values displayed in the report are reduced by the blank value.

