

Application Sheet

IMTEC-SmD1-Antibodies

REF ITC60029

Instrument: Elisis Duo

REF 17200

Setting:

SmD_QT_ITC60029_R4.asy

SmD_CO_ITC60029_R4.asy

For all essential product information, please refer to the User Manual of the analyzer and the instructions for use for reagents, controls, calibrators and the Certificate of Analysis. Please refer to the instructions for use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyzer.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

Material Required

Material	REF	Comments
IMTEC-SmD1-Antibodies	ITC60029	
Sample Rack Elisis Duo	17200/23	CAL, NC, PC sample collection tubes
Reagent Rack (2 Lanes) Elisis Duo	17200/25	CON, SUB, STOP
Adapter 100 ml Human Square Bottle	17200/30	DIL
Dilution Plate	17200/255	predilution of samples

Additional Notes

The concentration of the PC must be checked manually since the target values are batch specific. The concentration of the NC must also be checked by the user.

OD-values of 3.5 or above are outside the photometer's dynamic range. For this reason they are displayed as 4.0.