

Calibration Values / Kalibrationswerte / Valores de calibración / Valeurs de calibration

Test name @ Method	Reference material/method	Unit	Assigned Value	Relative expanded uncertainty of measurement	Coverage factor k*
Testname @ Methode	Referenzmaterial/-methode	Einheit	Zugewiesener Wert	relative erweiterte kombinierte Messunsicherheit	Erweiterungsfaktor k
Nombre de la prueba @ Método	Material/método de referencia	Unidad	Valor asignado	Incertidumbre ampliada relativa de medición	Factor de cobertura k*
Nom du test @ Méthode	Matériel/méthode de référence	Unité	Valeur assignée	Incertitude élargie relative de mesure	Coefficient de couverture k*
ACID PHOSPHATASE a-Naphthylphosphate, 37°C	Commercial calibrator	U/l µkat/l	16.3 0.27	11%	2.0
ALBUMIN Iiquicolor Bromcresol green	ERM-DA 470k/IFCC	g/dl g/l	5.25 52.5	4.2%	2.0
ALKALINE PHOSPHATASE Iiquicolor AMP buffer	IFCC reference method, 37°C	U/l µkat/l	284 4.73	2.8%	2.0
ALKALINE PHOSPHATASE opt. Iiquicolor DEA buffer, 37°C	Commercial method employing DEA buffer	U/l µkat/l	407 6.78	8.8%	2.0
alpha-AMYLASE Iiquicolor CNP3	IFCC reference method, 37°C	U/l µkat/l	280 4.67	3.0%	2.0
auto-BILIRUBIN-D Iiquicolor DPD method	Manual procedure with Bilirubin direkt Iiquicolor	mg/dl µmol/l	2.44 41.7	12.3%	2.0
auto-BILIRUBIN-T Iiquicolor DPD method	Doumas bilirubin reference method (re-optimized), 25°C	mg/dl µmol/l	4.03 69.0	3.1%	2.1
CALCIUM Iiquicolor Ortho-cresolphthalein	ICP-OES, NIST SRM915b	mg/dl mmol/l	12.2 3.05	1.6%	2.0
CHLORIDE Iiquicolor TPTZ	Coulometry, NIST SRM919b	mg/dl mmol/l	408 115	1.7%	2.3
CHOLESTEROL Iiquicolor CHOD-PAP	IDMS	mg/dl mmol/l	309 7.98	1.4%	2.5
CHOLINESTERASE Iiquicolor Butyrylthiocholine, DGKC 37°C	Manual procedure using molar extinction coefficient	U/l µkat/l	7906 132	3.7%	2.0
CK NAC IiquiUV Enzymatic	IFCC reference method, 37°C	U/l µkat/l	726 12.1	2.6%	2.0
auto-CREATININE Iiquicolor Jaffé	IDMS	mg/dl µmol/l	5.12 453	1.0%	2.6
CREATININE (enzym) Iiquicolor Enzymatic	IDMS	mg/dl µmol/l	5.12 453	1.0%	2.6
gamma-GT Iiquicolor Gamma-Glutamyl-3carboxy-4-nitroanilide	IFCC reference method, 37°C	U/l µkat/l	139 2.32	2.8%	2.0
GLUCOSE Iiquicolor GOD	IDMS, NIST SRM917c	mg/dl mmol/l	197 11.0	1.0%	2.0
GLUCOSE IiquiUV^{mono} Hexokinase	IDMS, NIST SRM917c	mg/dl mmol/l	197 11.0	1.0%	2.0
GOT (ASAT) IFCC mod. IiquiUV Enzymatic	IFCC reference method, 37°C	U/l µkat/l	122 2.04	2.5%	2.0
GPT (ALAT) IFCC mod. IiquiUV Enzymatic	IFCC reference method, 37°C	U/l µkat/l	119 1.98	2.3%	2.0
HDL CHOLESTEROL Iiquicolor Homogenous enzymatic assay	UC-IDMS	mg/dl mmol/l	139 3.60	3.7%	2.0
IRON Iiquicolor CAB	ID-ICP-MS	µg/dl µmol/l	314 56.2	18.5%	2.0
IRON TPTZ Iiquicolor TPTZ	ID-ICP-MS	µg/dl µmol/l	314 56.2	18.5%	2.0

Test name @ Method	Reference material/method	Unit	Assigned Value	Relative expanded uncertainty of measurement	Coverage factor k*
Testname @ Methode	Referenzmaterial/-methode	Einheit	Zugewiesener Wert	relative erweiterte kombinierte Messunsicherheit	Erweiterungsfaktor k
Nombre de la prueba @ Método	Material/método de referencia	Unidad	Valor asignado	Incertidumbre ampliada relativa de medición	Factor de cobertura k*
Nom du test @ Méthode	Matériel/méthode de référence	Unité	Valeur assignée	Incertitude élargie relative de mesure	Coefficient de couverture k*
LDH SCE mod. liquiUV Substrate Pyruvate, 37°C	Manual procedure using molar extinction coefficient	U/l µkat/l	666 11.1	13.1%	2.0
LDL CHOLESTEROL Iquicolor Homogenous enzymatic assay	UC-IDMS	mg/dl mmol/l	130 3.35	11.8%	2.0
LIPASE Iquicolor - from lot 21001 Substrate Methylresorufin, 37°C	Commercial calibrator	U/l µkat/l	72.5 1.21	10.3%	2.0
LIPASE Iquicolor - before lot 21001 Substrate Methylresorufin, 37°C	Commercial calibrator	U/l µkat/l	63.0 1.05	14.2%	2.0
MAGNESIUM Iquicolor Xylidyl blue	ICP-OES, NIST SRM929a	mg/dl mmol/l	3.04 1.25	1.9%	2.0
PANCREAS-AMYLASE Iquicolor EPS-G ₇	Manual procedure using molar extinction coefficient	U/l µkat/l	153 2.55	13.6%	2.0
PHOSPHORUS Iquirapid Molybdate (UV)	NIST SRM 3139a	mg/dl mmol/l	8.69 2.81	6.9%	2.0
TOTAL PROTEIN Iquicolor Biuret	NIST SRM927f	g/dl g/l	9.11 91.1	2.8%	2.2
TRIGLYCERIDES Iquicolor^{mono} GPO POD	IDMS	mg/dl mmol/l	239 2.73	1.2%	2.6
UREA IquiuV Urease (UV)	IDMS	mg/dl mmol/l	143 23.8	1.0%	2.6
URIC ACID Iquicolor^{plus} Uricase	IDMS	mg/dl µmol/l	10.1 598	1.9%	2.3
URIC ACID Iquicolor Uricase	IDMS	mg/dl µmol/l	10.1 598	1.9%	2.3

*The expanded uncertainty is declared and calculated by multiplying the standard uncertainty with the coverage factor *k*. The value of the measurable quantity lies within the attaches range