

Standard RHEUMATOID FACTORS

Version 1/10 - 2021

REF

11361

LOT

21504



31.03.2023

Calibration Values / Kalibrationswerte / Valores de calibración / Valeurs de calibration

Device Gerät Instrumento Equipement	Reference material/method Referenzmaterial/-methode Material/método de referencia Matériel/méthode de référence	Unit Einheit Unidad Unité	Assigned Value Zugewiesener Wert Valor asignado Valeur assignée
HUMAN RHEUMATOID FACTORS - Immunoturbidimetry			
Multipurpose reagent for other	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	150
		kIU/l	150
HumaStar System reagent on HumaStar 600 LOT ≤ 20003	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	150
		kIU/l	150
HUMAN RHEUMATOID FACTORS - Immunoturbidimetry Latex enhanced			
Multipurpose reagent on HumaStar 100 & 200	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	150
		kIU/l	150
Multipurpose reagent on HumaLyzer photometer	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	150
		kIU/l	150
Multipurpose reagent for other	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	150
		kIU/l	150
HumaStar System reagent on HumaStar 300SR LOT ≥ 21000	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	150
		kIU/l	150
HumaStar System reagent on HumaStar 600 LOT ≥ 21000	Adjusted to 1 st British Standard NIBSC 0064/2	IU/ml	150
		kIU/l	150

CA_RF_11361

Rev. 001 | valid of 17.03.2021

INF 1136102 - Version 1/10 - 2021

Human

Human Gesellschaft für Biochemica und Diagnostica mbH
Max-Planck-Ring 21 · 65205 Wiesbaden · Germany
Telefon +49 6122-9988-0 · Telefax +49 612-9988-100 · e-Mail human@human.de