


# PCT Control HumaCLIA SR

**REF** 85820

**LOT** 21002

Version 01 / 02 - 2022

Target Values / Sollwerte / Valores asignados / Valeurs cibles

 2022-12-10

Link for latest version of target value sheet:

<https://www.human.de/clia/target-value-sheets/>



<b>REF</b>	Assay Parameter Ensayo	Valid from Gültig ab Válido a partir de	Units Einheiten Unidades	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo
	Essai	Valable à partir de <b>RGT</b> Lot	Unité		ET	Intervalle		ET	Intervalle
				Level 1			Level 2		
85020	PCT	21001	ng/ml	0.43	0.04	0.31 - 0.55	10.54	1.05	7.39 - 13.69