


PCT Control HumaCLIA SR

REF 85820

LOT 22002

Version 02 / 11 - 2022

Target Values / Sollwerte / Valores asignados / Valeurs cibles

 2023-06-07

Link for latest version of target value sheet:

<https://www.human.de/clia/target-value-sheets/>



REF	Assay Parameter Ensayo	Valid from Gültig ab Válido a partir de	Units Einheiten Unidades	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo	Target Sollwert Valor asignado Cible	SD SD SD	Range Bereich Intervalo
	Essai	Valable à partir de RGT Lot	Unité	ET	Intervalle	ET	Intervalle		
				Level 1			Level 2		
85020	PCT	22001	ng/ml	0.60	0.09	0.33 - 0.87	12.92	1.46	8.54 - 17.30
85020	PCT	22002	ng/ml	0.60	0.09	0.34 - 0.86	13.47	1.35	9.56 - 17.39



Human